Adhesive Skin Collection Kit Instructions





Reminder: No thick hair or blood on samples

Ensure all required information is filled out and accurate on the Sample Collector Trifold. This information is required for the sample to be processed.



STEP 1: Trim excessive hair with scissors. Ensure no blood.



STEP 2: Fill out front of Sample Collector Trifold. All information is required for sample to be processed.



STEP 3: Thoroughly cleanse lesion site with enclosed 70% alcohol pad.



STEP 4:

Thoroughly dry by gently pressing enclosed gauze pad over lesion. Do not proceed to sample collection until lesion is completely dry.



STEP 5:

Open Sample Collector Trifold. Remove and discard blue paper liner and "Remove Before Use" seal.



STEP 6: Remove first Smart Sticker"



Hold skin tight and center first Smart Sticker directly over



STEP 8: Use thumb to firmly press Smart Sticker for 10 seconds. This will help maximize the amount of tissue collected.



STEP 13: Re-confirm patient information is completed and accurate on Sample Collector Trifold, Place sample into re-sealable foil bag containing desiccant pouch.



STEP 9: Use provided ultra-fine permanent marker to outline the mole. Ensure lesion is fully enclosed by the outline.

PLEASE NOTE:

- · Do not freeze or refrigerate collected sample.
- · Store at room temperature until shipment.
- Test results will be returned within 5 business days of recieivng the sample.



STEP 10: Remove the Smart Sticker by slowly and continuously peeling it from the skin in one direction.



STEP 11: Place the collected sample with adhesive side down directly on pre-marked oval area in Sample Collector Trifold

Sonora Quest Laboratories[™]

of Laboratory Sciences of Arizona



STEP 7:

lesion

STEP 12: Repeat steps 7-11 until all four (4) Smart Stickers have been used on the lesion.

